



PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidanan, Nabunturan, Davao de Oro

TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 25-2760

Bid Number: B-25-0143
Date: 04/22/2025
Page: 1 of 4

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
1	1.0	unit	<p>Clinical Centrifuge</p> <ul style="list-style-type: none"> • Approximately 2 seconds after the start of centrifugation it detects abnormal vibration and alerts that there is an imbalance. • User-friendly display. LCD display lights turn on and off repeatedly when the centrifugation is complete. • Three Short-Cut Program Keys for DirectAccess. • Motorized lid lock system. • A light touch is all that is needed to activate the lid motor and lock it automatically. • Mixed Loading where two different types of buckets can be centrifuged at the same time. Moreover, frequently used combinations of buckets are offered as a set as standard so that the accessories become more affordable. • Technical specifications: <ul style="list-style-type: none"> o Swing Rotor o Max capacity: 15mL x 24 o Max speed (rpm): 4,000rpm o Max RCF (xg): 2,610xg o Size: 34(W)x42(D)x33(H) cm Height with opened lid 66 cm o Weight: 20kg o Power consumption: 170W, 0.61 MJ/h o Calorific value: (Power-saving mode: Approx. 7W) o Power requirements: <ul style="list-style-type: none"> Single phase 110/115V ±10%, 50/60Hz, 8A Single phase 220/230V ±10%, 50/60Hz, 5A o Rated voltage, rated current: 110V 2.7A, 115V 2.5A, 220/230V 1.3A o Use environment: <ul style="list-style-type: none"> Temperature: 10-25°C Humidity: 30 - 85% Atmospheric pressure: 70 – 106kPa (700 – 1,060 mbar) o Conforming standard: C61010-2-020: 110/115V model CE Marking: 220/230V model o Speed setting range: 200-4,000 rpm 10 rpm step o Timer setting range: 	297,000.00	297,000.00		
SUB-TOTAL					297,000.00		
Purpose: FOR THE USE OF BLOOD AND LABORATORY 2ND QTR Delivery: PGSO Period of Delivery: 60 DAYS							



PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidanan, Nabunturan, Davao de Oro

TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 25-2760

Bid Number: B-25-0143
Date: 04/22/2025
Page: 2 of 4

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			Seconds: 1 second step Minutes: 1 minute step Hour: 1 hour step Hold o Program memory: 99 channels at the maximum (3 for direct invocation using the panel button) o Acceleration/deceleration: 2 ways of switching, natural deceleration Terms and Conditions 1. General Provisions 1.1 These terms and conditions ("Terms") govern the use, maintenance, and repair of laboratory equipment ("Equipment") both during the warranty period and after the expiration of the initial warranty. 1.2 By using the Equipment, the provincial government of Davao de Oro ("Customer") agrees to comply with these Terms. 2. Supplier's Responsibilities During Warranty Period 2.1 The Supplier/Distributor shall provide a warranty for the Equipment for one year starting from the date of purchase. 2.2 The Supplier/Distributor shall repair or replace defective parts caused by defects in materials or workmanship under normal use and maintenance at no cost to the Customer. 2.3 The Supplier/Distributor is responsible for processing warranty claims upon receipt of proof of purchase and a detailed description of the defect from the Customer. 2.4 During the warranty period, the Supplier/Distributor shall conduct regular monthly visits to check the Equipment's status and ensure its proper functioning. 2.5 The Supplier/Distributor is required to have a service center within the Davao Region to ensure immediate response in case of Equipment breakdown. 3. Warranty Exclusions 3.1 The Supplier/Distributor is not responsible for the following: a. Damage caused by improper use, neglect, or unauthorized modifications by the Customer. b. Routine maintenance and consumable items.				
GRAND TOTAL					297,000.00		
Purpose: FOR THE USE OF BLOOD AND LABORATORY 2ND QTR Delivery: PGSO Period of Delivery: 60 DAYS							



PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidanan, Nabunturan, Davao de Oro

TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 25-2760

Bid Number: B-25-0143
Date: 04/22/2025
Page: 3 of 4

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			<p>c. Damage resulting from external factors.</p> <p>3.2 The Supplier/Distributor will void the warranty if repairs or alterations are made by unauthorized personnel.</p> <p>4. Maintenance During Warranty</p> <p>4.1 The Supplier/Distributor shall provide preventive maintenance guidelines to the Customer to help maintain the warranty's validity.</p> <p>4.2 Preventive maintenance services may be performed by the Supplier/Distributor or an authorized service provider during the warranty period.</p> <p>5. Supplier's Responsibilities Post-Warranty</p> <p>5.1 After the warranty period, the Supplier/Distributor shall continue to offer repair and maintenance services on a chargeable basis.</p> <p>5.2 Upon receiving repair requests, the Supplier/Distributor shall provide an estimate of repair costs and obtain the Customer's approval before proceeding with any work.</p> <p>6. Replacement Parts</p> <p>6.1 The Supplier/Distributor shall supply genuine replacement parts for the Equipment upon request.</p> <p>6.2 The Supplier/Distributor shall ensure timely shipment of replacement parts, with associated costs borne by the Customer.</p> <p>7. Service Contracts</p> <p>7.1 The Supplier/Distributor shall offer optional service contracts for ongoing maintenance and support.</p> <p>7.2 These service contracts shall include various tiers with different service levels and response times to suit the Customer's needs.</p> <p>8. Technical Support</p> <p>8.1 The Supplier/Distributor shall provide technical support via phone, email, or onsite visits. Support may be subject to additional service fees as applicable.</p> <p>8.2 The Supplier/Distributor shall ensure a 24/7 technical assistance service to address urgent issues.</p> <p>8.3 The Supplier/Distributor is responsible for maintaining records of support requests and resolutions for</p>				
GRAND TOTAL					297,000.00		
<p>Purpose: FOR THE USE OF BLOOD AND LABORATORY 2ND QTR</p> <p>Delivery: PGSO</p> <p>Period of Delivery: 60 DAYS</p>							



PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidanan, Nabunturan, Davao de Oro

TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 25-2760

Bid Number: B-25-0143
Date: 04/22/2025
Page: 4 of 4

Table with columns: ITEM NO., QTY., UNIT OF ISSUE, DESCRIPTION, APPROVED BUDGET (Unit Price, Total Amount), BID PRICES (Unit Price, Total Amount). Includes a row for 'reference' with detailed terms and a 'GRAND TOTAL' row.

Purpose: FOR THE USE OF BLOOD AND LABORATORY 2ND QTR
Delivery: PGSO
Period of Delivery: 60 DAYS